



**SWAMI DAYANAND HOSPITAL
EAST DELHI MUNICIPAL CORPORATION
DILSHAD GARDEN, SHAHDARA
DELHI-110095**

No.SDNH/2017/ 623

Dated 09/02/17

CORRIGENDUM


Sub :- Regarding correction of typographical error of technical specification of ICU Ventilator at point no 9

In reference of NIT No. 14524 for purchase of ICU Ventilators and ICU due on 21.02.2017.

The following typographical error of technical specification of ICU ventilators with advanced mode of closed loop ventilation is hereby been corrected and should be read as :-

1. TV-20-200 ml should be read as TV-20-2000 ml


Medical Superintendent
Dr. M. L. JAIPAL
Medical Superintendent
SDN Hospital/EDMC


Director IT
Civic Centre