

**Tender Document for Purchase of Cardiac Monitors – 10 Nos.**

1. Tender Amount: Rs. 45.00 Lac
2. Earnest Money: 1,12,500/-

Schedule Display - Health Dept.

Tender No:	13923	Tender Ref Date:	27-06-2014
Tender Ref No:	No.010/281PMT/PC/2014/19dated26.06.2014		

**Schedule : Single Submission Multi Opening MCD**

Seq No	Health Dept. Stage	Contractor Stage	Start Date & Time	Expiry Date & Time	View Forms	Envelops
1	Release Tender	-	27-06-2014 10:00	04-07-2014 17:00	-	-
2	-	Purchase and Download Tender Document	04-07-2014 17:01	25-07-2014 17:00	✓	Financial Envelope, Technical Envelope
3	-	Online Bid Preparation and Hash Submission	04-07-2014 17:01	25-07-2014 17:00	✓	Financial Envelope, Technical Envelope
4	Technical and Financial Lock	-	25-07-2014 17:01	26-07-2014 17:00	-	Financial Envelope, Technical Envelope
5	-	Re-encryption of online bids	26-07-2014 17:01	28-07-2014 17:00	✓	Financial Envelope, Technical Envelope
6	Technical Bid Opening	-	28-07-2014 17:01	29-07-2014 17:00	-	Technical Envelope
7	Evaluation of Technical Bid	-	29-07-2014 17:01	02-08-2014 17:00	-	Technical Envelope
8	Financial Bid Opening	-	02-08-2014 17:01	03-08-2014 17:00	-	Financial Envelope
9	Evaluation of Financial Bid	-	03-08-2014 17:01	20-08-2014 17:00	-	Financial Envelope
10	Award of Contract	-	20-08-2014 17:01	20-09-2014 17:00	-	Financial Envelope, Technical Envelope

**Fee Structure**

SlNo	Stage	Formula
1	Purchase and Download Tender Document	4000

Man

**Technical Specification of Cardiac Monitors – 10 Nos.**

1. Monitor should be suitable for wall mounted fixed bedside patient monitoring.
2. Monitor should be suitable for Adult and Pediatric Patients.
3. Monitor should have minimum 15" size screen.
4. Each Monitor should have at least following standard Parameters:
  - a) ECG
  - b) Heat Rate (HR)
  - c) Non – Invasive Blood Pressure (NIBP)
  - d) Invasive Blood Pressure
  - e) Oxygen Saturation (SpO2) with Masimo SET
  - f) Respiration
  - g) Side stream ETCO2 – required in 05 Monitors only
5. Monitor should have minimum 6 waveforms of ECG, Plethysmography, Respiration, ETCO2 or more parameters.
6. Monitor should have different colors for different parameters and must be clearly visible from a distance of 10-12 feet. It should have Audio and visual alarms for the different parameters. The alarm should continue till manually switched off and the alarm limit should be easily adjustable.



**Dr. Mahindra Prasad**  
CMO / P&S

7. Monitor should be modular having separate slots for requisite modules / parameters – ETCO<sub>2</sub>, IBP etc. The different modules of ETCO<sub>2</sub> & IBP can be interchange between all the monitors.
8. Monitor should be connectable by wiring to Central Nursing Station.
9. ECG and HR Monitoring should have
  - a) Leads I, II, III, AVR, AVL, AVF and V (3/5 lead cable) or more
  - b) HR Range – 20-250 bpm or more
  - c) HR Accuracy +- bpm or +- 2% or better.
10. NIBP Monitoring should have
  - a) Operation Mode – Manual/ automatic/ Stat or more.
  - b) Adjustable time in Auto mode.
  - c) Must measure Systolic, Diastolic, Mean Pressures
  - d) Range – Systolic (60-250mm hg or more), mean (40-200mm Hg or more), Diastolic (30-180 mm Hg or more)
11. Invasive BP Monitoring should have:
  - a) Range --10 to300mm Hg or more
  - b) Accuracy - +-2mm Hg or +-2% or better.
12. SPO<sub>2</sub> Monitoring should have:
  - a) Range – 35 -100% or more
  - b) Accuracy - +- 3% or 3+- digits or +- 3 SD or better for 70-100%
13. Respiration Monitoring should have
  - a) Range - 8 -80 or more
  - b) Accuracy - +- 2 bpm or better
  - c) Apnea Alarm
14. Monitor should have at least a 24-hour Trend of parameters
15. Monitor should have minimum 6 display waveforms simultaneously.
16. Monitor should have Adjustable Upper and Lower Alarm limits with over-riding alarm.
17. Monitor should have a Printer Facility for central Station
18. Central Station screen size of 21" should have:
  - a) Acquire data from at least 8 or more monitors simultaneously
  - b) Minimum 2 selectable parameters of each patient.
  - c) Be able to be wired with bedside monitors.
19. Monitor should either have a built-in and re-chargeable Battery or a UPS should be provided for un-interrupted monitoring for two hours or more.
20. Electrical Requirement : 230 Volt AC, 50/60Hz
21. Operational Temperature : 10 C-35 C or more
22. Weight : less than 10 Kg

General Conditions.

1. A single price should be quoted to include 10 monitors with 10 sets of each module of requisite parameters including 05 Nos ETCO<sub>2</sub> modules, 10 Nos. 5-lead ECG Cables, 25 Nos. SPO<sub>2</sub> Probes, 5 Nos. Pediatric SPO<sub>2</sub> Probes, 15 Nos. adult NIBP cuffs, 5 Nos. pediatric NIBP cuffs, 10 Nos. IBP transducers and 10 sets of consumables for invasive IBP, 01 No. printer with 10 printer paper rolls for central station, 10 Nos. power cords, 10 Nos. rechargeable batteries and chargers / UPS and any other accessories / consumables required to make the equipment fully functional along with a Central Nursing Station and include fixing with wall mounts of monitors, as well as, their wiring and connections with the central station.
2. The company must provide a warranty for services, as well as, spares for 5 years to be valid from date of satisfactory installation and must quote rates for CMC charges of next 5 years to be effective after the completion of warranty.
3. Quoting firm must have a service centre in Delhi.
4. The company must submit a proof of the equipment being of national or international quality (FDA-USE/CE).
5. Proof of technical specifications should be clearly visible in the equipment's brochure or catalogue.
6. Compliance report should be mentioned point by point for all the specifications.
7. The monitor must remain functional at all the times and in case of breakdown, the monitor must be repaired within 24 hours or else this would be counted as down time and company will be required to extend warranty for carryout period.



**Dr. Mahindra Prasad**  
CMO / P&S

8. Rates to be quoted for item not a part of CMC.
9. Availability of spares of the equipment for 10 years post installations.

**Terms and conditions:**

1. Tenderers must agree to all terms and conditions of N-DMC, offers with counter terms and conditions are liable to be summarily rejected.
2. For livery and stationary items, tenderers must submit samples as per the given specifications along-with the offers.
3. Total quantity of purchase shown in the offer letter are only approximate and are liable to be increased or decreased according to the actual requirement of the institution.
4. No tender will be considered without earnest money. The earnest money will not be refunded till the supplies are completed satisfactorily
5. Director, RBIPMT reserves the right to reject any or all the tenders without assigning any reason.
6. Sealed and labeled samples of the articles, patent or own, with full information should be submitted. Wherever it is not possible self explanatory detailed Catalogue/brochure should be provided with the tender.
7. The approved rates of the firm for each item will be treated as rate valid for six months from the date of opening of the tenders/Negotiation.
8. The supply shall be made in full or in part according to the order placed within given stipulated period, failing which the tenderer shall be subjected to penalty of 1% on the total amount of order for every week delay to maximum 5% in case of default beyond that in addition to the penalty and the quotee shall be liable for difference in cost if purchase from other sources, entire risk and responsibility.
9. After the approval of item/samples (duly sealed and labeled), which the quotee shall supply, will have to be kept in the hospital for comparing with the future supplies.
10. All the deliveries will be quoted FOR RBIPMT, DELHI-110009, and all kinds of charges such as packing, delivery, transit railway freight, and demurrage and likewise should be clearly indicated wherever applicable.
11. The approved supplier shall be responsible to deliver goods in perfect, well and sound conditions at RBIPMT, Kingsway Camp, Delhi-110009. The supply is subject to approval of Director, RBIPMT/Board.
12. Supplies in case is considered inferior to or not in accordance with the approved samples, it shall be rejected. The quotee shall have to remove the articles with in a week at his own cost.
13. The successful tenderer will have to execute an agreement on non-judicial stamp paper of Rs. 50/- for supply and sign the necessary contract document with in seven days of the intimation of acceptance of the quotation. In case of delay earnest money will be forfeited. No grace of time period shall be allowed in any case.
14. No quotee shall converse with any official of N-DMC in respect of this quotation. Any quotee found indulging in such his activity will be blacklisted and his tender summarily rejected.



**Dr. Mahindra Prasad**  
CWO/P&S

15. In event of any dispute the decision of Director, RBIPMT shall be final and binding on the contractor.
16. The Director, RBIPMT shall have power to make any alteration, omission, addition or substitution as per hospital requirements if need be. The changes will be made in writing. Any such alteration, omission, addition or altered substitution shall not in any ways effect or invalidate the contract and tenderer may be directed to make supply in as part of work or supply.
17. Tender/quotation should be submitted along with DATASHEET and other relevant documents.
18. VAT Registration Copy has to be submitted.
19. Latest Clearance of VAT should be submitted.
20. Bidding documents can be downloaded from the website-<http://mcdetenders.com>
21. Both tender cost and earnest money can be paid through Net Banking or can be deposited in form of the Bank Draft in favour of Commissioner, NDMC of our scheduled Bank, at RBIPMT.
22. The desirous contractors shall have to pay the Tender Cost at the time of downloading of tender documents and earnest money at the time of Re-encryption of Online Bid.
23. If the payment is made through Demand Draft then the contractor has to deposit the demand draft for tender cost and earnest money before opening of Tender.
24. No previous paid earnest money will be adjusted.
25. Multiple offers will not be accepted.
26. No conditional offer/ quotation will be accepted. The contractor shall be responsible for submission of the fake information and the tender shall be rejected.
27. Drug License is required is to be submitted for medicines and drugs tenders.
28. 5% of contractual amount shall be deposited by the successful bidder in favour of Commissioner NDMC before the supply of material in shape of FDR as a performance guarantee/Security.
29. All supplies should be supplied with sufficient shelf life period.
30. All supplies should be mentioned "MCD SUPPLY NOT FOR SALE" on each unit of supply.
31. Form 39 / test report of Govt. approved Lab is required with supply of medicines.
32. Test analysis reports are parenteral preparations, surgical consumables (whichever is applicable) must compulsory declare **sterility, pyrogens & toxicity tests**. Similarly ophthalmic preparation wherever applicable should have sterility test.
33. Test analysis report of liquid preparations must also reflect the **volume** of the contents among other criteria.
34. Only manufacturer / their authorized agent will be considered for purchase. If the tender is submitted by the third person, he has to submit original copy of agreement with the manufacturer / authorized agent in the name of Director,



**Dr. Mahindra Prasad**  
CMO / P&S

RBIPMT, that the manufacturer / authorized agent will be responsible for guarantee/warranty of equipment along with

- a. The CMC along with availability of spares and consumables will also be responsibility of manufacturer / original vendor for total 10 years.
- b. The company should certify that equipment is latest and is not going to be obsolete in near future.
- c. Rate quoted are not more than those quoted in the other Govt. Institution.

35. The firm should be able to demonstrate the quoted equipment in working order preferably in NCR.

36. The equipment should be available for patient use continuously & un-interruptedly throughout the year.

37. The firm can also quote any other additional accessories if desired for the proper functioning of the equipment.

38. The firm should give warranty for 5 years and thereafter quote the CMC rates for the next 5 years, on year-wise basis.

39. The firm should also give a copy of CMC done in other Govt./Corporate Hospitals. A list of exhaustive consumable rates should be quoted (same as quoted in other Govt./Corporate Hospitals) by the manufacturer/sole agent/Authorized dealer.

40. The firm should provide certificate for availability of spare parts of equipment for 10 years.

41. The firm should have installation of the same equipment in the Govt. Institute / corporate institute in NCR

42. The firm should have satisfactory working equipment certificate from two other Govt./Corporate Hospitals.

43. The certificate regarding No Vigilance/CBI Inquiry against the firm has to be submitted.

44. The company should not be black listed/debarred and certificate in this regard is to be submitted by the firm.

45. The rates must be quoted inclusive of VAT. (applicable)

46. A certificate is required from the vendor/firms that the CMC/AMC shall be provided by the manufacturer/authorized distributor.

47. The vendor/firm has to comply with the order No. F.3(11)/Fin(Rev-I)/2012-13/dsVI/57-62 dated 17.01.2013 of department of finance, NCT of Delhi.

- a. "The goods, including for works contract, shall be supplied by bidder or its authorized distributor in Delhi and against a sale invoice issued from Delhi. The delivery of goods shall also be made from Delhi. The bidder dealer or its authorized distributor, as the case may be, who supplies the goods should be registered with the Delhi VAT Department and carry a valid tax identification Number issued by it. The bidder shall, however, be responsible for compliance with all conditions, warranties/guarantees, irrespective of the fact that the goods are supplied by him directly or through its authorized distributor. Further, the quoted bid price in the tender shall be inclusive of all taxes and duties."
- b. The Pre authorized agencies shall also ensure the following:  
"Delivery of goods is made from Delhi and against a sale invoice issued from Delhi. The dealer supplying goods should be registered with the Delhi VAT Department and carry a Valid Tax Identification Number issued by it"



**Dr. Mahindra Prasad**  
Chauras

48. The supply period for indigenous item will be 49 days and 90 days for imported items.

**TERMS AND CONDITIONS FOR FIRM:**

1. The firm quoting equipment should be a reputed firm
2. Only manufacturer / their authorized agent will be considered for purchase. If the tender is submitted by the third person, he has to submit original copy of agreement from the manufacturer / authorize agent in the name of Director, RBIPMT, that the manufacturer / authorized agent will be responsible for guarantee/warranty of equipment along with a. The CMC along with availability of spares and consumables, will also be responsibility of manufacturer / original vendor for total 10 years. b. The company should certify that equipment is latest and is not going to be obsolete in near future. c. Rate quoted are not more than quoted in the other Govt. Institute. d. Any other query from time to time in patient care only will be put up to the original vendor.
3. The firm should be able to demonstrate the equipment in working order preferably in NCR.
4. The equipment should be available for patient use continuous & un-interruptedly round the year.
5. The firm can also quote any other additional accessories if desired for the proper functioning of the equipment.
6. The firm should give quoted warranty for 5 years and thereafter CMC rates for the next 5 years, on year-wise basis.
7. The firm should be manufacturer / sole distributor of manufacturer.
8. The firm should also give a copy of CMC done in other Govt. Hospital list of exhaustive consumable rates should be quoted (same as quoted in other Govt. Hospitals.)
9. The firm should provide spare of equipment for 10 years.
10. The firm should have installation of the same equipment in the Govt. Institute / corporate institute in NCR.
11. The firm should have satisfactory working equipment certificate from other Govt. Hospitals.
12. The firm should be reputed and quote good quality latest equipment.

Director/RBIPMT

AMS/RBIPMT

CMO/P&S

PC/RBIPMT



**Dr. Mahindra Prasad**  
CMO P&S

**TENDER DOCUMENT:**

Tender Amount: Rs. 35.00 Lac  
 Earnest Money: Rs. 87500/-  
 Schedule Display - Health Dept.

Tender No: 13924 Tender Ref Date: 27-05-2014  
 Tender Ref No: No.DIR/RBIPMT/FC/2014/200dated26.05.2014

**Schedule : Single Submission Multi Opening MCD**

Seq No	Health Dept. Stage	Contractor Stage	Start Date & Time	Expiry Date & Time	View Forms	Envelops
1	Release Tender		27-06-2014 10:00	04-07-2014 17:00	-	
2		Purchase and Download Tender Document	04-07-2014 17:01	25-07-2014 17:00	✓	Financial Envelope, Technical Envelope
3		Online Bid Preparation and Hash Submission	04-07-2014 17:01	25-07-2014 17:00	✓	Financial Envelope, Technical Envelope
4	Technical and Financial Lock		25-07-2014 17:01	26-07-2014 17:00	-	Financial Envelope, Technical Envelope
5		Re-encryption of online bids	26-07-2014 17:01	28-07-2014 17:00	✓	Financial Envelope, Technical Envelope
6	Technical Bid Opening		28-07-2014 17:01	29-07-2014 17:00	-	Technical Envelope
7	Evaluation of Technical Bid		29-07-2014 17:01	02-08-2014 17:00	-	Technical Envelope
8	Financial Bid Opening		02-08-2014 17:01	03-08-2014 17:00	-	Financial Envelope
9	Evaluation of Financial Bid		03-08-2014 17:01	20-08-2014 17:00	-	Financial Envelope
10	Award of Contract		20-08-2014 17:01	20-09-2014 17:00	-	Financial Envelope, Technical Envelope

**Fee Structure**

SrNo	Stage	Formula
1	Purchase and Download Tender Document	1000

Man

**PURCHASE OF PLEURAVIDEOSCOPE (VIDEO THORACOSCOPE):  
WITH RIGID INSERTION TUBE AND DEFLECTABLE TIP**

- Light weight & fully immersible in disinfection solution.
- Provision for Autoclavability (preferably for all detachable components)
- Preferably 3 or 4 nos. of remote switches for user's convenience to control operation functions.
- In-built scope ID Memory chip.
- Compatible with semi automatic leakage tester with air flow regulation from attached light source air pump.
- Compatible with electrosurgical and laser treatment (YAG,810 nm Diode)
- Insulation (ceramic) tip for protection while doing electrosurgical procedure.
- Field of view : 120degree or more
- Depth of field : 3 mm to 100mm or better
- Direction of view : Forward viewing
- Distal End Dia. : 7.2 mm or less
- Insertion tube Dia. : 7.2 mm or less
- Working Length : 260 mm or more
- Min. Visible distance : 3 mm from distal End
- Instrument Channel Dia. : 2.8 mm or more
- Flexible tip bending range : Up-150 deg. Or more & Down-120 deg or more



**Dr. Mahindra Prasad**  
 CMO / P&S

**Standard set should include** following items otherwise to be quoted separately

- Swing jaws –Biopsy forceps(with needle) 10 Nos. (Swing Jaws, Oval Cup)
- Cytology Brushes (2 sets of minimum 10 pieces)
- Cleaning and maintenance kit
- Semi disposable leakage tester with its air flow & pressure regulation through air pump in compatible light source.
- Spray Catheter (1 no.)
- Electrosurgical coagulator.
- Flexible trocar set for safety of bending tube of thoracoscope.

**Video system center /video processor with keyboard should be**

- Compatible and latest Video-Processor for the equipment along with all the standards attachments necessary to make the equipment fully functional in every aspect must be provided.
- Video recording facility with all the hardware and compatible for data storage. Retrieval, procedure recording and report printing must be provided.
- Data entry software to be installed.
- Computer (Intel® Core™ 13-3220 Processor 3.30 GHz, 3MB Cache, with Hyper-Threading Technology, Windows 8 Single Language, English 64bit, 8GB Dual Channel 1600 MHz, 1 GB GDDR5 NVIDIA® GeForce® GTX, Slot-loading Dual Layer DVD Burner (DVD+RW, CD-RW, McAfee, LiveSafe/Symantec 12 months subscription & 1 year on-site warranty) with UPS and Color Laser printer to be provided.
- **Compatible Monitor:**
  - Compatible High resolution 17 inch. LCD colour monitor.
  - Compatible Online UPS for providing un-interrupted 60 minutes backup for the instrument.

**Xenon Light Source (300 watts):**

- Main Lamp :Xenon Short arc lamp(300watts) with switching Regulator mechanism
- Main Lamp life :Appx.500 hrs on continuous use.
- Capacity of producing special NBI light for future Upgradability.
- Emergency lamp: Halogen 12V 100 Watts (Life 100hrs or more)
- Automatic brightness & exposure control
- Air feeding Pump: With different levels of air feeding pressures (On,Off Low Medium ,High)
- Electric shock protection: Type Class I with Type Bf applied part.
- Power supply :220-240 V AC ,Frequency 50/60 Hz,

**Approximate cost –Rupees 35 lakhs**

Terms and conditions:

1. Tenderers must agree to all terms and conditions of N-DMC, offers with counter terms and conditions are liable to be summarily rejected.
2. For livery and stationary items, tenderers must submit samples as per the given specifications along-with the offers.
3. Total quantity of purchase shown in the offer letter are only approximate and are liable to be increased or decreased according to the actual requirement of the institution.
4. No tender will be considered without earnest money. The earnest money will not be refunded till the supplies are completed satisfactorily
5. Director, RBIPMT reserves the right to reject any or all the tenders without assigning any reason.
6. Sealed and labeled samples of the articles, patent or own, with full information should be submitted. Wherever it is not possible self explanatory detailed Catalogue/brochure should be provided with the tender.



**Dr. Mahindra Prasad**  
JIFS



7. The approved rates of the firm for each item will be treated as rate valid for six months from the date of opening of the tenders/Negotiation.
8. The supply shall be made in full or in part according to the order placed within given stipulated period, failing which the tenderer shall be subjected to penalty of 1% on the total amount of order for every week delay to maximum 5% in case of default beyond that in addition to the penalty and the quotee shall be liable for difference in cost if purchase from other sources, entire risk and responsibility.
9. After the approval of items/samples (duly sealed and labeled), which the quotee shall supply, will have to be kept in the hospital for comparing with the future supplies.
10. All the deliveries will be quoted FOR RBIPMT, DELHI-110009, and all kinds of charges such as packing, delivery, transit railway freight, and demurrage and likewise should be clearly indicated wherever applicable.
11. The approved supplier shall be responsible to deliver goods in perfect, well and sound conditions at RBIPMT, Kingsway Camp, Delhi-110009. The supply is subject to approval of Director, RBIPMT/Board.
12. Supplies in case is considered inferior to or not in accordance with the approved samples, it shall be rejected. The quotee shall have to remove the articles with in a week at his own cost.
13. The successful tenderer will have to execute an agreement on non-judicial stamp paper of Rs. 50/- for supply and sign the necessary contract document with in seven days of the intimation of acceptance of the quotation. In case of delay earnest money will be forfeited. No grace of time period shall be allowed in any case.
14. No quotee shall converse with any official of N-DMC in respect of this quotation. Any quotee found indulging in such his activity will be blacklisted and his tender summarily rejected.
15. In event of any dispute the decision of Director, RBIPMT shall be final and binding on the contractor.
16. The Director, RBIPMT shall have power to make any alteration, omission, addition or substitution as per hospital requirements if need be. The changes will be made in writing. Any such alteration, omission, addition or altered substitution shall not in any ways effect or invalidate the contract and tenderer may be directed to make supply in as part of work or supply.
17. Tender/quotation should be submitted along with DATASHEET and other relevant documents.
18. VAT Registration Copy has to be submitted.
19. Latest Clearance of VAT should be submitted.
20. Bidding documents can be downloaded from the website- <http://mcdetenders.com>
21. Both tender cost and earnest money can be paid through Net Banking or can be deposited in form of the Bank Draft in favour of Commissioner, NDMC of our scheduled Bank, at RBIPMT.
22. The desirous contractors shall have to pay the Tender Cost at the time of downloading of tender documents and earnest money at the time of Re-encryption of Online Bid.

  
**Dr. Mahindra Prasad**  
CMO / P&S

23. If the payment is made through Demand Draft then the contractor has to deposit the demand draft for tender cost and earnest money before opening of Tender.
24. No previous paid earnest money will be adjusted.
25. Multiple offers will not be accepted.
26. No conditional offer/ quotation will be accepted. The contractor shall be responsible for submission of the fake information and the tender shall be rejected.
27. Drug License is required is to be submitted for medicines and drugs tenders.
28. 5% of contractual amount shall be deposited by the successful bidder in favour of Commissioner NDMC before the supply of material in shape of FDR as a performance guarantee/Security.
29. All supplies should be supplied with sufficient shelf life period.
30. All supplies should be mentioned "MCD SUPPLY NOT FOR SALE" on each unit of supply.
31. Form 39 / test report of Govt. approved Lab is required with supply of medicines.
32. Test analysis reports are parenteral preparations, surgical consumables (whichever is applicable) must compulsory declare **sterility, pyrogans & toxicity tests**. Similarly ophthalmic preparation wherever applicable should have sterility test.
33. Test analysis report of liquid preparations must also reflect the **volume** of the contents among other criteria.
34. Only manufacturer / their authorized agent will be considered for purchase. If the tender is submitted by the third person, he has to submit original copy of agreement with the manufacturer / authorized agent in the name of Director, RBIPMT, that the manufacturer / authorized agent will be responsible for guarantee/warranty of equipment along with
- The CMC along with availability of spares and consumables will also be responsibility of manufacturer / original vendor for total 10 years.
  - The company should certify that equipment is latest and is not going to be obsolete in near future.
  - Rate quoted are not more than those quoted in the other Govt. Institution.
35. The firm should be able to demonstrate the quoted equipment in working order preferably in NCR.
36. The equipment should be available for patient use continuously & uninterrupted throughout the year.
37. The firm can also quote any other additional accessories if desired for the proper functioning of the equipment.
38. The firm should give warranty for 5 years and thereafter quote the CMC rates for the next 5 years, on year-wise basis.
39. The firm should also give a copy of CMC done in other Govt./Corporate Hospitals. A list of exhaustive consumable rates should be quoted (same as quoted in other Govt./Corporate Hospitals) by the manufacturer/sole agent/Authorized dealer.



Dr. Mahindra Prasad  
CMO / P&S

