



SOUTH DELHI MUNICIPAL CORPORATION
(Health Department)
Office of Director Hosp. Admn. (M&TB)
18th Floor, Dr. SPM Civic Center, Dr. JLN Marg,
New Delhi 110002; Tel 22326804

No. Addl. DHA/M&TB/SDMC/2018/50

Date: 11/12/18

Amendment

In continuation of Department's NIT, bearing no. Addl. DHA (M&TB)/16 dated 29/11/2018, and on receipt of representations from Vendors/prospective bidders, following amendments/changes in specifications are hereby notified for information of all. These are:

S.No	Specifications as per NIT	Amendment requested and recommended
1	2. X-Ray Tube, A dual focus rotating anode X-Ray Tube having following focal spots: <ul style="list-style-type: none">• Small Focal spot : 1.0• Large Focal Spot: 2.0 Anode Heat storage capacity : 140 KHU	2. X-Ray Tube: A dual focus rotating anode X-Ray Tube having following focal spots: <ul style="list-style-type: none">• Small Focal spot : 0.6 - 1.2• Large Focal Spot: 1.2-2.0 Anode Heat storage capacity : 140 KHU & above
2	3. Stand Tube stand and the Detector Stand should be integrated on same floor rail.	Stand Tube stand and the Detector Stand should be integrated on same floor rail or Separate
3	5. Flat Panel Detector: One No. Fixed Flat Panel Detector in detector stand. <ul style="list-style-type: none">✓ Size of detector should be minimum : 17" X 17"✓ X-ray Detection Method: Photodiode: Amorphous Silicon (a-Si) Scintillator: Cesium Iodine (CsI);✓ Pixel Pitch should be minimum: 130 micro mm.✓ Active Image Matrix more than or equal to: 3KX3K✓ A/D Conversion : 14 bits or more DQE: 70% or more	Pixel Pitch 150 micron or less. Active Image Matrix more than or equal to 2K x 2K. (Other details remain same)
4	B. Control Panel <ul style="list-style-type: none">• Anatomical programming up to 2016 program functioning which automatically according to the body part selection.• A two-step hand switch with dual accepted images.	<ul style="list-style-type: none">• Anatomical programming up to 216 Program functioning or more which automatically according to the body part selection.• A two-step hand switch.

The other details of specifications remain unchanged
This is for information of all.

S. P. Singh
11/12/18
Addl. DHA (M&TB)

Copy to:

1. Director IT to kindly upload in the website of SDMC & be also displayed as 'flash' also.
2. O/c

Copy for information.

1. PS to Commissioner for information of Commissioner;
2. Addl. Commissioner(Health) & CA cum FA;
3. ADC(health) cum CA cum FA;
4. DCA F&G;
5. DCA/Health